

Inspection Report

Report No.:QA-210825001

Customer	Cypress		
Model No.	Neb Tester	Serial No.	06182106
Equipment	Nebulizer Tester		
Procedure Used	QA-3-J28 Neb Tester Calibration Procedure		
Calibration Date	2021.08.25	Recommended Recal. Date	2023.08.24
Condition of Calibration	Temp. 27 °C	R.H. 50 %	
Standards Employed			
I.D. No.	MFG/Model No.	Equipment	Due Date
04270010-3853	NEW-FLOW/ 505SLM7-TOP SENSOR 4S	Pressure Gauge Flowmeter	2021/10/31
Traceability	Cal. Source / Report No.		Calibration Date
RM002114	IPE&KATRONIC / IA28P011		2019/10/31
Manager:	Calibrated by:		
<i>Xiao ai quan</i> 2021.8.25		<i>Dian min 8/25</i>	

Quantity values of above measurement standards used in this calibration are traced to those of

Inspection Report

Report No.:QA-210825001

Comparison Data :			
Flow rate :			
Measurement (L/min)	Reading (L/min)	Error (L/min)	Specification (L/min)
2	1.99	-0.01	
4	4.11	0.11	±0.5
6	6.20	0.20	
Pressure :			
Measurement (PSI)	Reading (PSI)	Error (PSI)	Specification (PSI)
5	5.89	0.89	
10	10.77	0.77	
15	15.64	0.64	±1
25	24.66	-0.34	